

# IHE-RO TSWF Call

Tuesday, January 21, 2025, 10:00am Eastern Time

## Attendees:

Bob Pekarek  
David Wikler  
Sanjay Bari  
Istvan Matyas  
Thomas Schwere

## Minutes:

- I. Call to order at 10:00am ET.
- II. Thomas to send around the proposal for “Time Zone Support”. Will be discussed in detail at the next meeting.
- III. Stub plan:
  - a. Stub plan was introduced because some vendors cannot store the full plan. Upon receiving the plan from a TPS, a stub plan is created which references the real plan through the existing Referenced RT Plan Sequence (300C,0002). The stub plan gets a new SOP Instance UID.
  - b. UPS and BDI will reference the stub plan. How does the TDD know that the plan provided in the UPS/BDI is a stub plan or a real plan?
    - i. Today this happens through configuration on the TDD and in case of a stub plan, the TDD retrieves the real plan from its local DB.
    - ii. An option was discussed to introduce an artificial AE Title for the TDD local DB. That artificial AE Title would be configured on the TMS which then could populate the plan retrieval information accordingly. In that case the TMS would basically have to know the Instance UID of the real plan only.
  - c. **Action Item**: Thomas to sketch a diagram for the artificial AE Title approach.
- IV. Meeting adjourned at 10:59am ET.

## Open **action items** from previous meetings:

Meeting	Action Item	Owner	Status
2024-08-27	Draft a CP to introduce “Time Zone Support” as an option to TDW-II.	Thomas	WIP
2024-10-22	Write a document about the stub plan topic. This could potentially result in a correction proposal for TDW-II	Sanjay	Open (2025-01-21)
2024-11-26	Draft a CP adding a note that the TMS may set Fraction Number in the Scheduled Processing Parameters Sequence to the Clinical Fraction Number at its own discretion.	Thomas	WIP
2024-11-26	Bring up the following topic at the next DICOM WG-07: Add treatment record references to the BDI to safely support the partial/continuation case (option “Retain Original Treatment Records”).	David	Open
2024-11-26	Add a note in 3.59.4.1 about handling of optional input information items: “To determine if the actor can perform the procedure safely, all items in the input information shall be referenced in the BDI and RT Plan.” (Depends also on previous action item)	Thomas	Open

